



Pre-Mac – Model SWP Water Purifier

www.pre-mac.com

Device Information

The Pre-Mac model SWP water purifier is a portable hand-pump water treatment device. According to the manufacturer, microbiological treatment consists of filtration and disinfection. The device consists of flexible inlet tubing containing a fine mesh screen which provides coarse filtration, outlet tubing, a hand pump, all connected to a disposable plastic cartridge containing activated carbon cloth and iodine resin. The activated carbon cloth provides filtration and adsorption the iodine resin provides disinfection. Operation of the hand pump draws water through the fine mesh screen on the inlet tubing and sends the water through the activated charcoal cloth and then the iodine resin in the cartridge. The resin is designed to impart an iodine residual (typically 2-4 mg/L) in treated water that provides additional disinfection. The manufacturer directs users to provide a minimum of 2 minutes contact time before drinking. A 4-minute contact time is directed when treating water at temperatures of 5° C or less. The manufacturer also offers an optional field test kit for measuring iodine residual in treated water.

Effectiveness Against Microbial Pathogens

No test data is available for the model SWP water purifier using the U.S. Environmental Protection Agency (USEPA) Guide Standard and Protocol for Testing Microbiological Water Purifiers (reference 1). However, an independent laboratory conducted testing using the USEPA Protocol on an earlier Pre-Mac model water purifier that is very similar in operation and treatment technology (reference 2). Therefore, these results are considered applicable to the model SWP water purifier. The results showed the earlier model consistently met the 6-log and 4-log bacteria and virus removal/inactivation minimum requirements. The device did not consistently meet the minimum 3-log protozoan cyst removal/inactivation requirement when challenged with Cryptosporidium oocysts. These test results suggest the Pre-Mac SWP would meet the minimum 6-log bacteria and 4-log virus removal/inactivation requirements when used according to directions. The results also suggest the Pre-Mac SWP would not meet the required minimum 3-log cyst removal/inactivation for *Cryptosporidium* oocysts. Other independent testing using protocols other than the USEPA Protocol verify the ability of the Pre-Mac SWP to provide at least a 6-log bacteria and 4-log virus removal/inactivation. Available testing data did not include Giardia cysts as a test organism. Based on general knowledge of the treatment technologies used in the SWP water purifier (activated carbon and iodine resin), the device would not consistently provide a 3-log removal or inactivation of Giardia cysts when used as directed. If the iodine resin is a pentaiodide (I₅) resin, the device would be capable of reducing Giardia cysts if contact time after passage through the device were extended to at least 40 minutes (reference 3). However, since there is no device-specific testing data available using

Giardia cysts and we do not know the exact composition of the iodine resin, we must consider the device ineffective against Giardia cysts. Additional treatment is, therefore, necessary to remove or inactivate cysts. Based on evaluation of available data and considering the data did not include device-specific testing using the USEPA Protocol, the Pre-Mac model SWP water purifier receives one $\sqrt{}$ for bacteria, and viruses, and one X for Giardia cysts and Cryptosporidium oocysts (for an explanation of the rating checks click here). The following table summarizes the device's expected effectiveness against microbial pathogens, evaluation rating, and the mechanism by which pathogens are removed or inactivated:

Table. Expected Performance Against Microbial Pathogens.

Microbial Pathogen Type	Expected Performance	Evaluation Rating	Inactivation/removal Mechanism
Bacteria	> 6-log	$\sqrt{}$	Iodine disinfection with some size exclusion and adsorption
Viruses	> 4-log	$\sqrt{}$	Iodine disinfection with some adsorption
Giardia cysts	Not Effective	X	Some size exclusion, adsorption and iodine disinfection
Cryptosporidium oocysts	Not Effective	X	Some size exclusion and adsorption

Production Rate and Capacity

Inherent to the production rate and capacity of filtration devices is the quality of the raw water source. The actual production rate and capacity is dependent on the user and raw water quality. The manufacturer's stated production rate is 200 ml/min. The stated capacity is 50-100L.

Cleaning, Replacement, End of Life Indicator

When pumping becomes difficult or 50L of water has passed through the device, the device must be disposed. The device is not capable of being cleaned or backwashed. Instructions recommend discarding the first 0.05L of treated water if the SWP water purifier is new.

Weight and Size

The dry weight of the device is 60 grams. Dimensions are 2 cm diameter x 13 cm length.



Cost

The Pre-Mac SWP water purifier is not sold at stores in the United States. The device is available through online ordering and at stores outside of the United States. The device costs approximately \$40.

Device Evaluation

Based on evaluation of available data, the Pre-Mac model SWP water purifier is expected to provide 6-log bacteria and 4-log virus removal or inactivation under most water quality conditions expected. The SWP water purifier will not consistently provide a 3-log Giardia cyst and Cryptosporidium oocyst removal or inactivation. Additional treatment such as filtration with a 1 µm absolute filter will be necessary to remove these protozoan cysts. Iodine resin disinfection is the primary mechanism of bacteria and virus inactivation. The iodine resin inactivates bacteria, viruses, and some Giardia cysts through direct contact with the resin as well as through the iodine residual the resin imparts to the water. The device will also provide some filtration and adsorption of bacteria, viruses, Giardia cysts, and Cryptosporidium oocysts due to the activated charcoal cloth. There is no indicator of process failure on a real-time basis and end of device useful life is based on filter clogging, or by the user keeping track of the volume of water purified. The iodine resin releases decreasing amounts of iodine as usage continues, but there are no instructions on when to dispose of the device based on iodine residuals if measured using the optional field test kit. Inherent to treatment devices using filtration is the likelihood of clogging and reduced device capacity when treating highly turbid water. The iodine resin and residual are not expected to cause any adverse health effects to healthy adults who have no preexisting thyroid conditions or sensitivity to iodine. This device is not recommended for use by pregnant women (concern for fetus), people with known hypersensitivity to iodine, people with a history (or family history) of thyroid disease, and people from areas with chronic iodine deficiency (reference 3). The iodine residual imparted by the resin can cause a medicinal taste and color the water. Iodine can be neutralized by adding ascorbic acid (Vitamin C) or sodium thiosulfate, which will improve the taste and color. Flavored drink mixes can mask the flavor. However, neutralizers and flavor aids should not be added until after recommend contact times are achieved.

<u>Advantages</u>

- Independent testing using the USEPA Protocol with a similar Pre-Mac device suggests the SWP water purifier will provide 6-log bacteria and 4-log virus removal or inactivation when treating most water quality conditions expected.
- Very small and lightweight.
- Very easy to use.
- No adverse health effects expected in healthy adults with no iodine sensitivity.

USACHPPM Water Supply Management Program Phone (410) 436-3919; Email <u>water.supply@apg.amedd.army.mil</u>



COTS Purifiers – Army Study Program, Project No. 31-MA-03E0-05

Disadvantages

- Not effective against *Giardia* and *Cryptosporidium*. Additional treatment is necessary.
- Not recommended for use by pregnant women or people with iodine sensitivity
- Can impart color and medicinal taste.

References

- 1. USEPA, Registration Division Office of Pesticide Program, Criteria and Standards Division Office of Drinking Water. (1987). *Guide Standard and Protocol for Testing Microbiological Water Purifiers*. Washington, D.C.
- 2. U.S. Army Biomedical Research & Development Laboratory. (1993). *Evaluation of the Medical Efficacy of the Pre-Mac Model FWP Individual Water Purifier for Treating Microbiological Contaminants in Water*. (USABRDL Technical Report 9204). Frederick, MD. Prepared by Shaub, S.A., Hargett, H.T., Sterling, C.R., and Marshall, M.M.
- 3. U.S. Army Center for Health Promotion and Preventive Medicine. (2005). *Technical Information Paper; Iodine Disinfection in the Use of Individual Water Purification Devices*, Aberdeen Proving Ground, MD.

